EU DECLARATION OF CONFORMITY

REGULATION ON MEDICAL DEVICES (EU) 2017/745, ANNEX IV

ORKLA WOUND CARE AB

SRN: SE-MF-000021041

This EU Declaration of Conformity is issued under the sole responsibility of Orkla Wound Care AB.

We hereby declare that the Medical Device products listed below conform to the Regulation on Medical Devices (EU) 2017/745.

Conformity assessment route:
The products are class I medical devices and have been CE marked in accordance with MDR 2017/745 Article 52 (7); Conformity assessment procedure of class I medical devices.

Basic UDI-DI:	7340210002002Z		
EMDN (CND) code:	M04010199	**GMDN code:**	44990
Intended purpose:	The products are intended to be used as a mechanical barrier and for absorption of exudates from minor and larger superficial wounds.		

REF	Name of product		Risk class
6036	Salvequick	Plastic Plasters Refill	I

Common Specifications applied: No applicable CS available

Solna 2024-08-06

Johanna Brinck

Johanna Brinck (Aug 6, 2024 18:33 GMT+2)

Johanna Brinck
Head of Regulatory and Quality
Person Responsible for Regulatory Compliance
Orkla Wound Care AB

Postadress/Postal address
Orkla Wound Care AB
Box 1336
SE-171 26 Solna
www.cederroth.com

Besöksadress/Visiting address
Svetsarvägen 15
Solna
Sweden
www.orklahealth.com/wound-care/

Telefon/Telephone
010 142 64 00
International
+46 (0)10 142 64 00

Orgnr
556765-1756
VAT No:
SE556765175601